

Chapter 1

History of Thyroid Ultrasound



Robert A. Levine and J. Woody Sistrunk

Abbreviations

AACE	American Association of Clinical Endocrinologists
AIUM	American Institute of Ultrasound Medicine
ATA	American Thyroid Association
ECNU	Endocrine Certification in Neck Ultrasound
MHz	Megahertz

R.A. Levine, MD, FACE, ECNU (✉)
Geisel School of Medicine at Dartmouth College, Thyroid Center
of New Hampshire, St. Joseph Hospital, Nashua, NH, USA
e-mail: thyroidmd2@gmail.com

J.W. Sistrunk, MD, FACE, ECNU
Jackson Thyroid & Endocrine Clinic, PLLC, Jackson, MS, USA

© Springer International Publishing AG 2018

D.S. Duick et al. (eds.), *Thyroid and Parathyroid Ultrasound
and Ultrasound-Guided FNA*,

https://doi.org/10.1007/978-3-319-67238-0_1

Introduction

The visual application of sound in medicine has revolutionized the diagnosis and management of thyroid disease. The safety of ultrasound, along with improvements in image quality and equipment availability, underlies the importance of thyroid ultrasound to today's endocrinologist and endocrine surgeon.

The thyroid is amenable to ultrasound study because of its superficial location, vascularity, size, and echogenicity [1]. In addition, the thyroid has a very high incidence of nodular disease, the vast majority benign. Most structural abnormalities of the thyroid need evaluation and monitoring but may not require intervention [2]. Between 1965 and 1970, there were seven articles published specific to thyroid ultrasound. In the last 5 years, there have been over 10,000 articles published. Thyroid ultrasound has undergone a dramatic transformation from the cryptic deflections on an oscilloscope produced in A-mode scanning, to barely recognizable B-mode images, followed by initial low-resolution gray scale, to current high-resolution images. Recent advances in technology, including harmonic imaging, spatial compound imaging, elastography, and three-dimensional reconstruction, have all furthered the field.

The development of high-resolution thyroid ultrasound required decades of study in both the acoustics of sound and data processing. Some animals, for example, dolphins and bats, have the ability use ultrasound in their daily activities in everything from catching prey to finding a mate. As early as the 1700s, the Italian biologist Lazzaro Spallanzani demonstrated that bats use high frequency sound waves to navigate in complete darkness [3]. The aim of this chapter is provide an overview of the basic advancements in the field of ultrasound that have provided the ability to easily and safely see and interpret structures inside the neck.

Beginnings of Ultrasound History

One of the earliest experiments regarding transmission of sound was performed in 1826 in Lake Geneva by Jean-Daniel Colladon. Using an underwater bell he determined the speed of sound transmission in water. In the 1800s, properties of sound including wave transmission, propagation, reflection, and refraction were defined. In 1877 Lord Rayleigh's English treatise, "Theory of Sound," added mathematics and became the basis for the applied study of sound. The principles described lead to the science of using reflected sound in identifying and locating objects. In 1880, Pierre and Jacques Curie discovered the piezoelectric effect, determining that an electric current applied across a crystal would result in a vibration that would generate sound waves and that sound waves striking a crystal would, in turn, produce an electric voltage. Piezoelectric transducers were capable of producing sonic waves in the audible range and ultrasonic waves above the range of human hearing [3].

Sonar

The first patent for a sonar device was issued to Lewis Richardson, an English meteorologist, only 1 month after the Titanic sank following collision with an iceberg. The first functional sonar system was made in the United States, by Canadian Reginald Fessenden, in 1915. The Fessenden "fathometer" could detect an iceberg 2 miles away. As electronics improved, Paul Langevin designed a device called a hydrophone. It became one of the first measures available to detect German U-Boats during World War I. The hydrophone was the basis of the pulse-echo sonar that is still employed in ultrasound equipment today [3, 4].

Rudimentary high frequency ultrasound analysis was used on a commercial basis in the 1930s and 1940s to detect defects

in steel such as the hull of a ship. Although crude by today's standards, inhomogeneity suggested abnormalities, whereas a flawless appearance suggested uniform material [4]. With the end of World War II, the development of the computer and the invention of the transistor advanced the development of medical ultrasound [3].

Early Medical Applications of Ultrasound

The initial use of ultrasound in medicine in the 1940s was therapeutic rather than diagnostic. Following the observation that very high-intensity sound waves had the ability to damage tissues, lower intensities were tried for therapeutic uses. Focused sound waves were used to mildly heat tissue for therapy of rheumatoid arthritis, and early attempts were made to destroy the basal ganglia to treat Parkinson's disease [4]. The American Institute of Ultrasound in Medicine (AIUM) was formed in 1952 with therapeutic ultrasound in physical medicine being the primary focus. Although members performing diagnostic ultrasound were not accepted until 1964, diagnostic ultrasound is currently the primary focus of this organization [3].

Early in the twentieth century, Paul Langevin described the ability of high-intensity ultrasound to induce pain in a hand placed in a water tank. The 1940s saw therapeutic ultrasound tried in numerous applications ranging from gastric ulcers to arthritis. Attempts to destroy the basal ganglia in patients with Parkinson's disease now seem archaic. At the time therapeutic ultrasound was headed toward the museum of medical quackery, consideration of ultrasound as a diagnostic tool in medicine had begun. Although Drs. Gohr and Wedekindt at the Medical University of Koln, Germany, suggested that ultrasound could detect tumors, exudates, and abscesses, the results were not convincing. Karl Theodore Dussik is credited as the first physician to use diagnostic ultrasound. In his 1952 report, "Hyperphonography of the Brain," ultrasound was utilized in localizing brain tumors and

the cerebral ventricles by transmitting ultrasonic sound through the skull. While the results of these studies were later discredited as predominantly artifact, this work played a significant role in stimulating research into the diagnostic capabilities of ultrasound [3].

A-Mode Ultrasound

One of the first studies of diagnostic ultrasound was performed by George Ludwig. Using A-mode ultrasound, his main focus was using ultrasound to detect gallstones, shown as reflected sound waves on an oscilloscope screen. Through his study of various tissues, including the use of live subjects, clinical utility of diagnostic ultrasound was described. Despite the limited efficacy of his rudimentary ultrasound system, Ludwig's most important achievement may be his determination of the velocity of sound transmission in animal soft tissues. Ludwig also determined that the optimum frequency of an ultrasound transducer for deep tissue was between 1 and 2.5 MHz. The ultrasound characteristics of mammalian tissue were further defined by physicist Richard Bolt at Massachusetts Institute of Technology and neurosurgeon H. Thomas Ballantine, Jr. at Massachusetts General Hospital [3].

Most of early ultrasound used a transmission technique, but by the mid-1950s that was supplanted by a reflection technique. Providing information limited to a single dimension, A-mode scanning showed deflections on an oscilloscope indicating distance to reflective surfaces [4] (see Fig. 2.7). A-mode ultrasonography was used for detection of brain tumors, shifts in the midline structures of the brain, localization of foreign bodies in the eye, and detection of detached retinas [4]. In the first presage that ultrasound may assist in the detection of cancer, John Julian Wild reported the observation that gastric malignancies were more echogenic than normal gastric tissue. Along with Dr. John Reid, he later studied 117 breast nodules using a 15 MHz sound source and

reported the ability to determine their size with an accuracy of 90% [3].

B-Mode Ultrasound

During the late 1950s, the first two-dimensional B-mode scanners were developed. B-mode scanners display a compilation of sequential A-mode images to create a two-dimensional image (see Fig. 2.8). Douglass Howry developed an immersion tank B-mode ultrasound system which was featured in the Medicine section of Life Magazine in September 1954 [3]. Several additional models of immersion tank scanners followed. All utilized a mechanically driven transducer that would sweep through an arc, with an image reconstructed to demonstrate the full sweep. Continued development led to the “Pan-scanner,” a more advanced B-mode device, but it still employed a cumbersome bathtub of water. Later advances included a handheld transducer that still required a mechanical connection to the unit to provide data regarding location and water-bag coupling devices to eliminate the need for immersion [3].

By 1964, the work of Joseph H. Holmes along with William Wright and Ralph (Edward) Meyerdirk lead to the prototype of the “compound contact” scanner, with direct contact of the transducer with the patient’s body. As stated in a 1958 Lancet article describing ultrasound evaluation of abdominal masses, “Any new technique becomes more attractive if its clinical usefulness can be demonstrated without harm, indignity or discomfort to the patient” [5].

Applying Ultrasound Technology to the Thyroid

The 1960s brought continued development of microelectronics including semiconductors that revolutionized the ability to process signals and produce visual displays. The phased

array transducer utilized in modern day ultrasound derived from highly classified submarine technology. During the 1970s additional advances in transducer design, including the linear array and mechanical oscillating transducers, lead to the two-dimensional imaging which remains the standard today. With these improvements and the addition of gray-scale displays, ultrasound representation of the thyroid began to resemble that seen in the operative field or gross anatomy lab [4].

In 1967 Fujimoto reported data on 184 patients studied with a B-mode ultrasound “tomogram” utilizing a water bath [6]. The authors reported that no internal echoes were generated by the thyroid in patients with normal thyroid function and non-palpable thyroid glands. They described several basic patterns generated by palpably abnormal thyroid tissue. Thyroid tissue with strong internal echo attenuation characteristics was considered “malignant.” Unfortunately, 25% of benign adenomas showed the malignant pattern, and 25% of papillary carcinomas were found to have the benign pattern. Although the first major publication of thyroid ultrasound attempted to establish the ability to determine malignant potential, the results were nonspecific in a large percentage of the cases. However, this was a seminal paper in ultrasound and is considered the first on thyroid ultrasound to attempt to establish the malignant appearance of nodules [4, 6].

In 1971 Manfred Blum published a series of A-mode ultrasounds of thyroid nodules (see Fig. 2.7). He demonstrated the ability of ultrasound to distinguish solid from cystic nodules, as well as accuracy in measurement of the dimensions of thyroid nodules [7]. Additional publications in the early 1970s further confirmed the capacity for both A-mode and B-mode ultrasound to differentiate solid from cystic lesions but consistently demonstrated that ultrasound was unable to distinguish malignant from benign solid lesions with acceptable accuracy [8].

The advent of gray-scale display resulted in images that were far easier to view and interpret [6]. In 1974 Ernest Crocker published *The Gray Scale Echographic Appearance*

of Thyroid Malignancy. Using an 8 MHz transducer with a 0.5 mm resolution, he described “low amplitude, sparse and disordered echoes” characteristic of thyroid cancer when viewed with a gray-scale display [9]. The pattern felt to be characteristic of malignancy was what would now be considered “hypoechoic and heterogeneous”.

With each advance in technology, interest was rekindled in ultrasound’s ability to distinguish benign from malignant lesions. Initial reports of ultrasonic features typically described findings as being diagnostically specific. Later, reports followed showing overlap between various disease processes. For example, following an initial report that the “halo sign,” a rim of hypoechoic signal surrounding a solid thyroid nodule, was seen only in benign lesions [10], Propper reported that two of ten patients with this finding had carcinoma [11]. As discussed in Chap. 7, the halo sign is still considered to be one of the numerous features that can be used in determining the likelihood of malignancy in a nodule.

In 1977 Walfish recommended combining fine-needle aspiration biopsy with ultrasound in order to improve the accuracy of specimen acquisition [12]. Subsequent studies demonstrated that biopsy accuracy is greatly improved when ultrasound is used to guide needle placement. Most patients with prior “nondiagnostic” biopsies will have an adequate specimen obtained when ultrasound-guided biopsy is performed [13]. Ultrasound-guided fine-needle aspiration results in improved sensitivity and specificity, as well as a greater than 50% reduction in nondiagnostic and false-negative biopsies [14].

Over the past several years, the value of ultrasound in screening for suspicious lymph nodes prior to surgery in patients with biopsy proven cancer has been established. Current guidelines for the management of thyroid cancer indicate a pivotal role for ultrasound in monitoring for locoregional recurrence [15].

During the 1980s Doppler ultrasound was introduced, allowing detection of blood flow in tissues. As discussed in detail in Chap. 3, the role of Doppler in assessing the

likelihood of malignancy has undergone a recent reevaluation. Doppler imaging may demonstrate the increased blood flow characteristic of Graves' disease [16] and may be useful in distinguishing between Graves' disease and thyroiditis, especially in pregnant patients or when radioisotope scanning is unavailable (see Chap. 3). Doppler imaging is useful in determining the subtype of amiodarone-induced thyrotoxicosis [17].

Recent Advances in Technology

Recent technological advancements include intravenous sonographic contrast agents, three-dimensional ultrasound imaging, and elastography. Intravenous sonographic contrast agents are available in Europe but remain experimental in the United States. All ultrasound contrast agents consist of microspheres, which function both by reflecting ultrasonic waves and, at higher signal power, by reverberating and generating harmonics of the incident wave. Ultrasound contrast agents have been predominantly used to visualize large blood vessels and have shown promise in imaging peripheral vasculature as well as liver tumors and metastases [18]. While no studies have been published demonstrating any advantage of contrast agents in routine thyroid imaging, the use of contrast agents or B-flow imaging may be helpful in the immediate assessment of successful laser or radiofrequency ablation of thyroid nodules [19].

Three-dimensional display of reconstructed images has been available for CT scan and MRI for many years and has demonstrated practical application. While three-dimensional ultrasound has gained popularity for fetal imaging, its role in diagnostic neck ultrasound remains unclear. Obstetrical ultrasound has the great advantage of the target being surrounded by a natural fluid interface, greatly improving surface rendering, whereas 3D thyroid ultrasound is limited by the lack of a similar interface distinguishing the thyroid from adjacent neck tissues. It has been predicted that breast

biopsies may eventually be guided in a more precise fashion by real time 3D imaging [20], and it is possible that, in time, thyroid biopsy will similarly benefit. At present, however, 3D ultrasound technology does not provide a demonstrable advantage in thyroid imaging.

Elastography is a promising technique in which the compressibility of a nodule is assessed by ultrasound, while external pressure is applied. With studies showing good predictive value for detection of malignancy in breast nodules, recent investigations of its role in thyroid imaging have been promising. Additional prospective trials are ongoing to assess the role of elastography in predicting the likelihood of thyroid malignancy. The role of elastography in the selection of nodules for biopsy or surgery is discussed in Chap. 16.

Application of Neck Ultrasound by Endocrinologists and Endocrine Surgeons

With the growing recognition that real time ultrasound performed by a clinician provides far more useful information than that obtained from a radiology report, point of care ultrasound has gained acceptance. The first educational course specific to thyroid ultrasound was offered in 1998 by the American Association of Clinical Endocrinologists (AACE). Under the direction of Dr. H. Jack Baskin, 53 endocrinologists were taught to perform diagnostic ultrasound and ultrasound-guided fine-needle aspiration biopsy. By the turn of the century, 300 endocrinologists had been trained. Endocrine University, established in 2002 by AACE, began providing instruction in thyroid ultrasound and biopsy to all graduating endocrine fellows. By 2016 over 6000 endocrinologists had completed an AACE ultrasound course. In 2007 a collaborative effort between the American Institute of Ultrasound in Medicine (AIUM) and AACE established a certification program for endocrinologists trained in neck ultrasound. By 2016 the ECNU (Endocrine Certification in Neck Ultrasound) program had certified over 470 endocrinologists as having the

training, experience, and expertise needed to perform thyroid and parathyroid ultrasound and fine-needle aspiration biopsy. In 2011 the American Institute of Ultrasound in Medicine began accrediting qualified endocrine practices as centers of excellence in thyroid and parathyroid imaging. To date, 89 practices have received AIUM site accreditation in thyroid and parathyroid ultrasound.

Conclusion

When the American Association of Clinical Endocrinologists began its efforts to teach thyroid ultrasound to Endocrinologists in 1998, an ultrasound machine seemed a foreign concept in the office. At present, it is becoming the exception to find endocrinologists who do not have thyroid ultrasound and ultrasound-guided FNA biopsy as part of their practice.

In parallel with the growth of thyroid ultrasound in endocrinology, the American Thyroid Association (ATA) guidelines for the management of thyroid nodules and thyroid cancer have placed an increasing emphasis on the sonographic characteristics of thyroid nodules. The 2006 guidelines mention ultrasound characteristics of thyroid nodules five times [21]. The 2009 ATA guidelines make 14 references to ultrasound characteristics [22], and the latest 2015 ATA guidelines mention ultrasound characteristics of thyroid nodules and thyroid cancer 100 times [15].

In the 50 years since ultrasound was first used for thyroid imaging, there has been a profound improvement in the technology and quality of images. The transition from A-mode to B-mode to gray-scale images was accompanied by dramatic improvements in clarity and interpretability of images. Current high-resolution images are able to identify virtually all lesions of clinical significance. Ultrasound characteristics can predict which nodules are likely to be benign and detect features including irregular margins, microcalcifications, and central vascularity that may deem a nodule suspicious [4]. Ultrasound

plays a clear fundamental role in thyroid nodule and lymph node evaluation as well as the selection of which should undergo biopsy [15]. Ultrasound has proven utility in the detection of recurrent thyroid cancer in patients with negative whole body iodine scan or undetectable thyroglobulin [15, 23]. Recent advances including the use of contrast agents, tissue harmonic imaging, elastography, and multiplanar reconstruction of images have further enhanced the diagnostic value of ultrasound images. Ultrasound guidance of fine-needle aspiration biopsy has been demonstrated to improve both diagnostic yield and accuracy and has become the standard of care. Routine point of care use of ultrasound is often considered an extension of the physical examination by endocrinologists and endocrine surgeons. High-quality ultrasound systems are now available at prices that make this technology accessible to virtually all providers of endocrine care [4].

Acknowledgment The authors wish to acknowledge the work of Dr. Joseph Woo and his excellent web-based overview of the history of ultrasound. Relevant parts of his work with application to thyroid ultrasound have been presented here. For his full text, please access <http://www.ob-ultrasound.net/history1.html>.

References

1. Solbiati L, Osti V, Cova L, Tonolini M. Ultrasound of the thyroid, parathyroid glands and neck lymph nodes. *Eur Radiol.* 2001;11(12):2411–24.
2. Tessler FN, Tublin ME. Thyroid sonography: current applications and future directions. *AJR.* 1999;173:437–43.
3. Woo JSK. A short history of the development of ultrasound in obstetrics and gynecology. <http://www.ob-ultrasound.net/history1.html>. Accessed 29 June 2016.
4. Levine RA. Something old and something new: a brief history of thyroid ultrasound technology. *Endocr Pract.* 2004;10(3):227–33.
5. Donald I, Macvicar J, Brown TG. Investigation of abdominal masses by pulsed ultrasound. *Lancet.* 1958;271:1188–95.
6. Fujimoto F, Oka A, Omoto R, Hirsoe M. Ultrasound scanning of the thyroid gland as a new diagnostic approach. *Ultrasonics.* 1967;5:177–80.

7. Blum M, Weiss B, Hernberg J. Evaluation of thyroid nodules by A-mode echography. *Radiology*. 1971;101:651–6.
8. Scheible W, Leopold GR, Woo VL, Gosink BB. High resolution real-time ultrasonography of thyroid nodules. *Radiology*. 1979;133:413–7.
9. Crocker EF, McLaughlin AF, Kossoff G, Jellins J. The gray scale echographic appearance of thyroid malignancy. *J Clin Ultrasound*. 1974;2(4):305–6.
10. Hassani SN, Bard RL. Evaluation of solid thyroid neoplasms by gray scale and real time ultrasonography: the “halo” sign. *Ultrasound Med*. 1977;4:323.
11. Propper RA, Skolnick ML, Weinstein BJ, Dekker A. The nonspecificity of the thyroid halo sign. *J Clin Ultrasound*. 1980;8:129–32.
12. Walfish PG, Hazani E, Strawbridge HTG, et al. Combined ultrasound and needle aspiration cytology in the assessment and management of hypofunctioning thyroid nodule. *Ann Intern Med*. 1977;87(3):270–4.
13. Gharib H. Fine-needle aspiration biopsy of thyroid nodules: advantages, limitations, and effect. *Mayo Clin Proc*. 1994;69:44–9.
14. Danese D, Sciacchitano S, Farsetti A, Andreoli M, Pontecorvi A. Diagnostic accuracy of conventional versus sonography guided fine-needle aspiration biopsy in the management of nonpalpable and palpable thyroid nodules. *Thyroid*. 1998;8:511–5.
15. Haugen BR, Alexander EK, Bible KC, Doherty G, et al. 2015 American Thyroid Association management guidelines for adult patients with thyroid nodules and differentiated thyroid cancer. *Thyroid*. 2016;26(1):1–133.
16. Ralls PW, Mayekowa DS, Lee KP, et al. Color-flow Doppler sonography in Graves’ disease: “thyroid inferno.”. *AJR*. 1988;150:781–4.
17. Bogazzi F, Bartelena L, Brogioni S, et al. Color flow Doppler sonography rapidly differentiates type I and type II amiodarone induced thyrotoxicosis. *Thyroid*. 1997;7(4):541–5.
18. Grant EG. Sonographic contrast agents in vascular imaging. *Semin Ultrasound CT MR*. 2001;22(1):25–41.
19. Andrioli M, Valcavi R. Ultrasound B-flow imaging in the evaluation of thermal ablation of thyroid nodules. *Endocrine*. 2015;48(3):1013–5.
20. Lees W. Ultrasound imaging in three and four dimensions. *Semin Ultrasound CT MR*. 2001;22(1):85–105.

21. Cooper DS, Doherty GM, Haugen BR, Kloos RT, et al. Management guidelines for patients with thyroid nodules and differentiated thyroid cancer. *Thyroid*. 2006;16(2):109–42.
22. Cooper DS, Doherty GM, Haugen BR, Kloos RT, et al. Revised American Thyroid Association management guidelines for patients with thyroid nodules and differentiated thyroid cancer. *Thyroid*. 2009;19(11):1167–214.
23. Antonelli A, Miccoli P, Ferdeghini M. Role of neck ultrasonography in the follow-up of patients operated on for thyroid cancer. *Thyroid*. 1995;5(1):25–8.