

- Fullerton, CA: Society of Critical Care Medicine, 1981, vol 2:1-53
19. Falke KJ, Pontoppidan H, Kumar A, et al. Ventilation with end-expiratory pressure in acute lung disease. *J Clin Invest* 1972;51:2315-2323
 20. Jonmarker C, Jansson L, Honson B, et al. Measurement of functional residual capacity by sulfur hexafluoride washout. *Anesthesiology* 1985;63:89-95
 21. Jonmarker C, Castor R, Drefeldt B, Werner O. An analyzer for in-line measurement of expiratory sulfur hexafluoride concentration. *Anesthesiology* 1985;63:84-88
 22. Webster JG. Medical instrumentation, application and design. Boston: Houghton-Mifflin, 1978:342-344
 23. Oppenheim AV, Schaffer RW. Digital signal processing. Englewood Cliffs, NJ: Prentice-Hall, 1975:26-30
 24. Armengol JA, Jones RL, King EG. Comparison of plethysmographic methods for obtaining thoracic gas volume in anesthetized dogs. *J Appl Physiol* 1985;59:653-657
 25. Fletcher R, Werner O, Nordström L, Jonson B. Sources of error and their correction in the measurement of carbon dioxide elimination using the Siemens-Elema CO₂ analyzer. *Br J Anaesth* 1983;55:177-185

CORRECTION

In "An Electroencephalographic Processing Algorithm Specifically Intended for Analysis of Cerebral Electrical Activity" by Thomas K. Gregory and Daniel C. Pettus (July 1986), an error appeared in the legend for Figure 4. The plot of amplitude response is not in decibels, as indicated in the legend. The first sentence of the legend should have read as follows: "Plots of amplitude response versus frequency (in Hertz) for (a) the two-pass 30-Hz filter and (b) the two-pass 30-Hz filter followed by the two-pass 8-Hz filter." We apologize for the error.