

ARTICLE

Open Access

# In-situ diagnostic of femtosecond laser probe pulses for high resolution ultrafast imaging

Chen Xie<sup>1,2</sup>, Remi Meyer<sup>2</sup>, Luc Froehly<sup>2</sup>, Remo Giust<sup>2</sup> and Francois Courvoisier<sup>2</sup>

## Abstract

Ultrafast imaging is essential in physics and chemistry to investigate the femtosecond dynamics of nonuniform samples or of phenomena with strong spatial variations. It relies on observing the phenomena induced by an ultrashort laser pump pulse using an ultrashort probe pulse at a later time. Recent years have seen the emergence of very successful ultrafast imaging techniques of single non-reproducible events with extremely high frame rate, based on wavelength or spatial frequency encoding. However, further progress in ultrafast imaging towards high spatial resolution is hampered by the lack of characterization of weak probe beams. For pump–probe experiments realized within solids or liquids, because of the difference in group velocities between pump and probe, the determination of the absolute pump–probe delay depends on the sample position. In addition, pulse-front tilt is a widespread issue, unacceptable for ultrafast imaging, but which is conventionally very difficult to evaluate for the low-intensity probe pulses. Here we show that a pump-induced micro-grating generated from the electronic Kerr effect provides a detailed in-situ characterization of a weak probe pulse. It allows solving the two issues of absolute pump–probe delay determination and pulse-front tilt detection. Our approach is valid whatever the transparent medium with non-negligible Kerr index, whatever the probe pulse polarization and wavelength. Because it is nondestructive and fast to perform, this in-situ probe diagnostic can be repeated to calibrate experimental conditions, particularly in the case where complex wavelength, spatial frequency or polarization encoding is used. We anticipate that this technique will enable previously inaccessible spatiotemporal imaging in a number of fields of ultrafast science at the micro- and nanoscale.

## Introduction

The fundamental understanding of laser matter interaction in several fields of ultrafast physics and chemistry requires imaging with both high spatial resolution (typ. sub-1  $\mu\text{m}$ ), and high temporal resolution (typ. sub-100 fs). This is the case for instance for laser wakefield acceleration<sup>1</sup>, amplification in laser-excited dielectrics<sup>2</sup>, ultrafast ionization and plasma formation<sup>3</sup>, THz radiation<sup>4,5</sup>, high harmonic generation<sup>6</sup>, new material synthesis via laser-induced microexplosion<sup>7</sup> or laser nanoscale processing<sup>8,9</sup>. The initial concepts of ultrafast imaging based on

repetitive pump–probe measurements<sup>10–15</sup> have been recently complemented by a large number of different schemes allowing the imaging of non-reproducible events. This is performed via a sophisticated probe sequence or compressed photography<sup>16–24</sup>, where the temporal information is encoded in the probe wavelength and/or in the spatial spectrum.

However, further progress in ultrafast imaging at high resolution is still impeded by two problems. First, a key information is the absolute delay between pump and probe. This is crucial to link the excitation dynamics to the actual pump pulse intensity<sup>25</sup>. Although synchronizing pump and probe pulses at a sample surface seems reasonably easy, the case of the synchronization of pump and probes in the bulk of a sample remains unaddressed. The problem is particularly acute when bulky microscope objectives impose pump and probe beams to pass through

Correspondence: Francois Courvoisier (francois.courvoisier@femto-st.fr)

<sup>1</sup>Ultrafast Laser Laboratory, Key Laboratory of Opto-electronic Information Technology of Ministry of Education, School of Precision Instruments and Opto-electronics Engineering, Tianjin University, 300072 Tianjin, China

<sup>2</sup>FEMTO-ST Institute, Univ. Bourgogne Franche-Comté, CNRS, 15B avenue des Montboucons, 25030 Besançon Cedex, France

These authors contributed equally: Chen Xie, Remi Meyer

© The Author(s) 2021



**Open Access** This article is licensed under a Creative Commons Attribution 4.0 International License, which permits use, sharing, adaptation, distribution and reproduction in any medium or format, as long as you give appropriate credit to the original author(s) and the source, provide a link to the Creative Commons license, and indicate if changes were made. The images or other third party material in this article are included in the article's Creative Commons license, unless indicated otherwise in a credit line to the material. If material is not included in the article's Creative Commons license and your intended use is not permitted by statutory regulation or exceeds the permitted use, you will need to obtain permission directly from the copyright holder. To view a copy of this license, visit <http://creativecommons.org/licenses/by/4.0/>.

the same optic, in a nearly collinear geometry. In this case, a sample longitudinal shift by only 100 micrometers, such as the one needed to image through a 150  $\mu\text{m}$  microscope glass slide, shifts the relative delay between colinear 800 nm pump and 400 nm probe pulses by 40 fs because of the difference between their group velocities in the dielectric material. In other words, the absolute delay is intrinsically bound to the exact position of the focus inside the bulk of the solid or liquid medium under study, and to the dispersive linear and nonlinear properties of the medium. The pump–probe absolute delay must be determined using a pump–probe interaction on a scale of a few tens of micrometers to obtain a temporal accuracy in pump–probe measurement below the 10 fs scale. Unfortunately, conventional pulse synchronization techniques are inoperable in this context. Sum frequency generation and other nonlinear frequency mixing schemes require high intensities in the probe pulse. Frequency mixing can be operated only in specific crystals and with only a limited number of probe wavelengths. Polarization gating<sup>26–29</sup> usually requires several 100's  $\mu\text{m}$  to several mm of pump–probe overlap distance for phase accumulation<sup>30,31</sup>. Transient-Grating cross-correlation frequency resolved optical gating (TG-XFROG) technique was successfully used to measure low energy probe pulses from a supercontinuum<sup>32</sup> or even in the UV<sup>33</sup>. However, the conventional configuration is non-phase matched and the phase matching is only reached when the interacting waves have a sufficiently wide angular spectrum, i.e., when they are focused, which is incompatible with the ultrafast imaging techniques mentioned above<sup>16–23</sup>. Optical Kerr Effect cross-correlation was used in a spectral interferometry setup<sup>34</sup> which is again incompatible with imaging. Generating a high density plasma with the pump and imaging it in shadowgraphy-like measurements is also often used to synchronize pump and probes, but the technique is destructive and the peak density reached by the plasma does not necessarily correspond to the pump pulse peak at the sub-100 fs scale. The intensity-dependent focal shifts due to the Kerr effect make this type of measurement even more difficult to exploit, especially in the parallel pump–probe configuration.


A second particularly difficult issue in ultrafast imaging with high spatial resolution is the removal of pulse front tilt due to the angular dispersion of the probe pulse. This problem arises because of the necessary dispersion compensation of the temporal dispersion induced by microscope objectives<sup>35</sup>. (As an example, a  $\times 50$  microscope objective induces a dispersion in excess of  $\sim 12,000 \text{ fs}^2$  at 400 nm, which stretches a 70 fs pulse to nearly 500 fs.) Most of the compensation schemes rely on spatially spreading the pulse spectrum such that unavoidable small misalignment creates angular dispersion<sup>36</sup>. This can be usually neglected for very low numerical aperture. In

contrast, in the case of high resolution imaging, the magnification of the setup also multiplies the coefficient of angular dispersion<sup>37</sup>.

As we will see below, after a  $\times 50$  microscope objective, the pulse front tilt can easily exceed  $70^\circ$  (equivalent to a delay of 90 fs in 10  $\mu\text{m}$  field-of-view) for a misalignment of only 10 mrad in a prism compressor. Pulse front tilt can be measured using frequency conversion autocorrelation in various spatial schemes<sup>38–41</sup> or using spatio-spectral interferometry<sup>42,43</sup> which are both unpractical for low-intensity pulses at arbitrary central wavelengths with  $< 100 \mu\text{m}$  beam diameter. Pulse front tilt is commonly regarded as very difficult to evaluate, particularly for low-intensity broadband pulses such as those used in the recent ultrafast photography techniques based on wavelength encoding.

Here we demonstrate a highly sensitive in-situ diagnostic for weak probe pulses which solves these two issues in the bulk of dielectrics and is applicable to a large number of ultrafast imaging scenarios. The concept is shown in Fig. 1a and will be detailed in the next section. The pump pulse, shaped by a Spatial Light Modulator, creates a micrometric transient grating via the optical Kerr effect oriented in the Bragg condition for the probe beam. Transient gratings were used in the past for various applications, such as measuring spatial coherence<sup>44</sup>, coherence time<sup>45,46</sup>, or spectroscopy<sup>47,48</sup>. In contrast, the transient grating used here is of micrometric size and maximizes the signal using the Bragg orientation. The micrometric size of the transient grating (typ. 10–30  $\mu\text{m}$ , with a period of typ. 1.3  $\mu\text{m}$ ) generates a pump–probe interaction that is highly localized which also allows for preserving a uniform diffraction efficiency over the broadband spectra of ultrashort probes. The diffracted signal provides a localized characterization of the absolute pump–probe delay. In addition, we have designed a way to straightforwardly visualize the angular dispersion using temporal stretching of the probe, so as to efficiently remove pulse front tilt.

Our diagnostic is valid whatever the probe wavelength and polarization and uses intensities that are sufficiently low to avoid optical breakdown or damage. We have used probe pulse energy down to pJ level in the multishot integration regime. Therefore, this diagnostic can be repeated as many times as required, for instance at each sample replacement, to enable highly reproducible pump–probe imaging experiments. Here, the imaging is performed with 0.8 numerical aperture at 400 nm wavelength, corresponding to an Abbe resolution of 240 nm and temporal precision of 6 fs. After performing the diagnostic, the phase mask applied on the Spatial Light Modulator can be easily switched to generate the desired pump beam, such that this diagnostic can be adapted to a very large number of pump to probe orientations.



The paper is organized as follows. We first derive the diffracted signal and present the optical setup. We then demonstrate that it can be used whatever the polarization configuration. We experimentally demonstrate that we can retrieve the absolute pump–probe delay when the sample is longitudinally displaced and that the delay variation actually follows the difference of the group velocities between pump and probe. Last, we solve the second issue of pulse front tilt removal using a visualization tool based on pulse temporal stretching and observation of the diffracted signal in the far-field.

## Results

### Cross-correlation signal and setup

We form a two-wave interference field inside a dielectric sample (fused silica, sapphire, or glass) from a single pump beam, using a single Spatial Light Modulator, which automatically ensures the synchronization between the two pump waves. The instantaneous electronic Kerr effect

transforms the interference intensity pattern into a grating with a period  $\Lambda$  (see Fig. 1b). We rotate the transient grating by an angle  $\alpha$  to match the Bragg incidence condition for a probe pulse which is a collimated beam propagating along the optical axis. The rotation is simply performed by adding the same tilt angle  $\alpha$  on the two interfering pump beams using the SLM. Typically, here the crossing half-angle of the pump is  $\theta = 12^\circ$  and the rotation angle is  $\alpha = 6^\circ$ . The ratio between these two angles is simply the ratio between the probe and pump wavelengths (see “Methods” section). The interference pattern extends typically over a propagation distance of  $40\ \mu\text{m}$  (see Suppl. Material Fig. S1) and can be easily reduced to below  $10\ \mu\text{m}$  in length. Using the SLM, this was performed by reducing the diameter of the interfering pump beamlets. We note that we have experimentally tested a wide set of configurations with  $\theta$  ranging from  $5$  to  $25^\circ$  and obtained quantitatively the same synchronization results.

In the “Methods” section, we derive the diffraction efficiency, based on the coupled wave theory for thick gratings. The diffracted signal is a cross-correlation between the squared pump intensity and the probe intensity. The intensity in the first diffraction order, for a pump–probe delay  $\tau$ , reads:

$$I_{1\text{st}}(\tau) \propto \left(\frac{n_2}{\cos \alpha}\right)^2 \int I_{\text{pump}}^2(t) I_{\text{probe}}(t - \tau) dt \quad (1)$$

with  $n_2$  the Kerr index related to the relative polarizations states of pump and probe pulses.

This signal will be key to characterize the probe. The high-intensity pump pulse can be independently characterized with another technique such as self-referencing spectral interferometry<sup>49,50</sup>, or SHG- or self-diffraction-FROG techniques<sup>51</sup>. We note that the interaction is based on three plane waves, in contrast with conventional TG-XFROG where the phase matching is reached by crossing focused beams<sup>32</sup>. Preserving a probe beam as close as possible to a plane wave is important for further use in pump–probe imaging. The rotation of the interference field of the pump can be adapted to match the Bragg angle for any probe incidence, for instance to meet the requirements of ultrafast imaging with structured illumination<sup>21</sup>. In addition, this Bragg angle can be adjusted to any incident probe wavelength when encoding is based on the probe central wavelength in a very wide spectrum<sup>16</sup>. We note that the micrometric size of the Bragg grating is a benefit in terms of spectral reflectivity: in our experimental conditions, the diffraction efficiency peak width exceeds 50 nm FWHM, i.e., much larger than the bandwidth of a 10 fs pulse centered at 395 nm (see “Methods”).

Our experimental setup is described in Fig. 1c and detailed in the “Methods” section. We use a Ti:Sapphire chirped pulse amplifier (CPA) laser source which delivers ~50 fs pulses at 790 nm central wavelength and all measurements are performed by integrating the signal over 50 shots at 1 kHz repetition rate.

We split the beam in a pump and a probe, which is frequency doubled with a  $\beta$ -barium-borate (BBO) crystal generating 60 fs pulses full-width at half maximum (FWHM). The pump pulse is then spectrally filtered to reduce its bandwidth to 12 nm FWHM, avoiding chromatic dispersion in the beam shaping stage (the filter was chosen to avoid generating sub-pulses, see “Methods”). We spatially shape the pump beam using a spatial light modulator (SLM) in near-normal incidence. The illuminated SLM is imaged by a lens and microscope objective in a 2f-2f telescope configuration with a de-magnification factor of 208. The image of the SLM is positioned at the front focal plane of the microscope objective, within the sample. The pump pulse duration has been characterized to be ~115 fs at sample site, after the first microscope


objective. This element is, in contrast, highly dispersive for the probe beam at 395 nm central wavelength (on the order of ~12000 fs<sup>2</sup>). We, therefore, compensate the linear dispersion on the 395 nm probe with a folded two-prisms compressor<sup>36</sup>. The probe beam is also de-magnified by a factor of 278 so that the probe beam has a waist of 12  $\mu\text{m}$  (Rayleigh range of 1.1 mm) in the sample. The polarization state of the pump and probe pulses are independently controlled by the rotation of half-waveplates.

Our samples are here fused silica and sapphire, whose Kerr index is  $n_2 \simeq 3 \times 10^{-16} \text{ cm}^2 \text{ W}^{-1}$  at a wavelength of 800 nm<sup>52,53</sup>. Their electronic Kerr index is instantaneous (see “Methods”). After interaction in the sample, we collect the pulses with a second  $\times 50$  (N.A. 0.8) microscope objective (MO). A relay lens images the Fourier plane of the second microscope objective onto a camera. The camera consequently records the far-field of the diffracted beams. In the k-space, the different orders of diffraction are separated. Figure S1 in Supplementary Material shows the characterization of the generated pump interference field and of the unperturbed probe beam.

### A Kerr-based transient grating valid for all combinations of pump–probe polarizations

Here, we validate that the measured diffracted signal effectively follows Eq. (1) and demonstrate that the measurement is valid for all combinations of input pump and probe polarizations. We remark that for sake of clarity, we report first on the validation of our technique and second, we report on the optimization of the probe pulse. In practice, we first removed the pulse front tilt due to the prism compressor. Then, we optimized the duration of the probe pulse, before recording the data used for the validation shown in Fig. 2. Here, the probe pulse duration is 60 fs. In all figures below, the pump–probe delay is positive when the probe arrives after the pump pulse at the position of the transient grating in the sample.

Figure 2a shows in the inset the recorded cross-correlation signal measured on the camera as a function of the relative pump–probe delay, for different pump pulse energies, indicated on top of each curve. Here, pump and probe pulses have the same horizontal polarization state. We observe that all curves have identical profiles, peaked at the same position for energies above 0.2  $\mu\text{J}$  (more comments are provided in section 3 of the Suppl. Mat). The main Fig. 2a shows the evolution of the peak signal as a function of the pump pulse energy. It fits very well with a quadratic curve of the input pump energy as expected from Eq. (1). The measurements have been performed in glass (Schott D263 microscope glass slide) and the results were also reproduced with identical conclusions in sapphire. Therefore, at this input power level ( $10^{11}$  to  $10^{13} \text{ W cm}^{-2}$ ), no contribution from higher order nonlinearities or electron plasma formation is



observable. The peak intensity remains low because the interfering pump beams are in a collimated configuration (waist of  $\sim 10\ \mu\text{m}$ , see Fig. S1).

In Fig. 2b, we show the evolution of the cross-correlation signal for the four different linear polarization configurations: both pump and probe can be either horizontally or vertically polarized. The grating period is oriented vertically, as shown in Fig. 1a.

The effective Kerr index  $n_2$  depends on the relative direction between pump and probe polarizations<sup>54,55</sup>. Indeed, for an isotropic medium like glass,  $n_{2//} = 3n_{2\perp}$  where  $n_{2//}$  is the Kerr index when pump and probe polarizations are parallel, and  $n_{2\perp}$  corresponds to the case where these polarizations are orthogonal to each other. Therefore, the signal efficiencies between parallel polarizations and orthogonal polarizations follow the ratio  $(\frac{n_{2//}}{n_{2\perp}})^2 = 3^2$ . In our measurements, the signal ratio is in a range 6–10 when varying the grating period  $\Lambda$ . This ratio is highly sensitive to the background subtraction. Despite the relatively large error bar, the experimental ratio is in very good agreement with the expected one.

These results overall confirm that the transient grating signal is effectively generated by Kerr effect. We obtain the pump–probe synchronization using the barycenter of the curve. The cross-correlation curve also straightforwardly allows the measurement of the compression of the probe pulse while tuning the prism compressor. This is shown in Suppl. Fig. S4. It provides a direct evidence of the optimal compression for the probe at the sample site. In our case, the cross-correlation curve allows us to retrieve the probe pulse duration of  $\approx 60\ \text{fs}$  FWHM knowing the pump pulse duration of  $100\ \text{fs}$  with  $2300\ \text{fs}^2$  of second order dispersion (see Suppl. Mat.). We note that


this in-situ diagnostic is also particularly useful when the sample itself is highly dispersive.

Finally, it is important to note that the technique is adaptable to characterize both polarizations. This is very useful to detect spectral phase differences in the optical path of the pump and probe beams. In Fig. 2b, in all four polarization cases, the signal is effectively peaked at the same delay whatever the combination of input pulses polarizations. However, in preliminary experiments, a non-optimal dichroic filter used to recombine pump and probe had a different spectral reflectivity for vertical and horizontal pump polarizations, as shown in Supplementary Fig. S3. For the horizontal pump polarization, it was inducing a spectral phase distortion. Our technique has identified this bias: a temporal shift as high as  $100\ \text{fs}$  and profile distortion was apparent from the cross-correlation curve. This highlights the effectiveness of the diagnostic even for the pump pulse.

### Spatial confinement of the synchronization

Since pump and probe pulses usually have different group velocities in the sample, the synchronization criterion, i.e., the absolute zero pump–probe delay, has to be defined for a precise location of the focus in the sample. In contrast with other synchronization or pulse measurement techniques, here, the interaction region between pump and probe is highly localized, down to some tens of micrometers. We successfully determined the pump–probe synchronization even for a transient grating length below  $10\ \mu\text{m}$ , obviously compromising on higher integration time to maintain an acceptable signal-to-noise ratio. The length of the transient grating can be adjusted using the SLM.





When we shift the position of the transient grating within the sample, we observe that the cross-correlation curve is shifted in delay. Experimentally, a first cross-correlation curve (Fig. 3a (top curve)) is acquired for a transient grating position starting at 50  $\mu\text{m}$  from the entrance surface of a 400  $\mu\text{m}$  thick sapphire sample, with refractive index  $n_g^{790} = 1.75$ . (The impact of the anisotropy of C-cut sapphire is negligible on beam shaping and on our measurements). When the sample is then shifted upstream by a distance  $d = 50 \mu\text{m}$ , the fringe pattern is shifted downstream by  $d(n_g^{790} - 1) = 37 \mu\text{m}$  because of refraction. When we repeat the cross-correlation measurement for different positions of the transient grating, we observe the cross-correlation shift in delay by 13 fs every 50  $\mu\text{m}$  shift. This corresponds to the difference in group velocities between 790 and 395 nm wavelength. Analytically, the delay induced by the group velocity difference between red and blue pulses is (see “Methods”):

$$\Delta t = (n_g^{790} - n_g^{395}) \frac{dn_g^{790}}{c} \quad (2)$$


where  $n_g^{395} = 1.796$  and  $n_g^{790} = 1.750$  are the group indices of sapphire at the central wavelengths of 395 and 790 nm<sup>56</sup>. We plot this curve in Fig. 3b as a dashed line and see that it perfectly fits with the experimental data of the position of the barycenter of the cross-correlation curves reported from Fig. 3a. Similarly, we obtained an excellent agreement in Fused silica

(see Suppl. Fig. S5), where the temporal delay is 22.6 fs every 100  $\mu\text{m}$  longitudinal shift. In microscope glass, the same shift induces a delay as high as 37 fs. We, therefore, demonstrate here that the strong localization of our measurement allows for retrieving the effect of the difference in group velocities on the pump–probe synchronization.

#### Diagnostic for the pulse front tilt of the probe beam

A prism compressor is aberration-free only when the two prisms are perfectly parallel. But it is experimentally unavoidable that the parallelism deviates by several milliradians. This deviation has however a dramatic impact on the probe pulse since it generates pulse front tilt, which is highly detrimental for the imaging of ultrafast phenomena. We will see here that the transient grating offers a straightforward visualization of the pulse front tilt. It can then be effectively canceled with the correct adjustment of the parallelism between the compressor prisms.

To evaluate how critical the problem is, we have evaluated the impact of a deviation angle from perfect parallelism between the two prisms in the prism compressor using ZEMAX<sup>TM</sup> software (see “Methods” section). After the prism compressor, for a deviation angle of 10 mrad of the second prism and compensation of the pointing angle with the folding mirror, the angular dispersion is 0.045 mrad nm<sup>-1</sup> with a negligible spatial chirp. However, the telescope used afterwards to decrease the probe beam



**Fig. 4 Visualization of angular dispersion.** **a** Concept of the diffraction of an angularly dispersed probe pulse by the transient grating. The transient grating effectively samples the chirped pulse at the pump–probe delay and diffracts the corresponding sub-pulse on the ROI (region of interest) in the first order of diffraction. **b** Typical experimental result. Diffracted signal as a function of delay and deviation angle in  $y$  direction.

waist to  $12\ \mu\text{m}$ , increases the angular dispersion by the inverse of the magnification<sup>37</sup>, i.e., a factor of 278. Quantitatively, at the focus of the microscope objective, the angular dispersion becomes  $12\ \text{mrad nm}^{-1}$ . Overall, a positioning error of only  $10\ \text{mrad}$  generates a significant pulse front tilt as high as 78 degrees which would dramatically blur the dynamics of ultrafast phenomena imaged. (This corresponds to a temporal delay exceeding 150 fs between the two sides of a  $10\ \mu\text{m}$  wide field of view). In the following, we will see how the transient grating can be used to detect and remove this strong pulse front tilt.

To detect angular dispersion and remove pulse front tilt, we develop a technique based on the fact that angular dispersion acts as a spectrometer. It spreads the spectral content of the probe pulse on the horizontal  $y$  axis, which corresponds to the direction of the angular mismatch in our prism compressor. Since the camera is placed in the Fourier space, each direction  $k_y$  is mapped onto a single column of pixels. When the probe pulse is temporally chirped, the transient grating samples the probe pulse spectrum in time. This is sketched as a concept in Fig. 4a: the different wavelengths are sampled by the transient grating at different moments (because of temporal chirp) and are diffracted into different directions (because of angular dispersion). Figure 4b shows a set of experimental images of the first diffracted order at different pump–probe delays when the probe pulse is slightly away from the optimal temporal compression. We observe the lateral shift of the diffracted spot along  $y$  direction with the pump–probe delay, similarly as in the concept Fig. 4a. In Fig. 5, we show the diffracted intensity as a function of  $k_y$  (converted in wavelength by angular dispersion) and pump–probe delay for different values of angular dispersion (adjusted with prism angle) and temporal dispersion  $\phi_2$  (adjusted with prism insertion). A single trace

is obtained by summing for each delay the diffracted signal shown in 4b along  $x$ -direction. Because of the angular dispersion acting as a spectrometer, these maps are similar to TG-XFROG signal maps, except this time in a phase-matched configuration because of the Bragg orientation.


When the spectral phase of the probe pulse is purely of second order, we have analytically demonstrated, in the Suppl. Material, that the signal in  $(k_y, t)$  space appears as an ellipse. We have derived the slope of its major axis in the limit of high chirp and small angular deviation. The slope of the major axis of the ellipse expresses as (see Suppl. materials):

$$\frac{dt}{dk_y} = \frac{\phi_2}{p} \quad (3)$$

where  $\phi_2$  is the second order dispersion and  $p$  the pulse front tilt<sup>57</sup>.

The top row of Fig. 5 shows the signal traces for a dispersion  $\phi_2 \simeq 8.5 \times 10^3\ \text{fs}^2$  at the sample site. The influence of the angular dispersion, and the corresponding pulse front tilt  $p$ , is highly apparent on the orientation of the major axis of the ellipse, which has been traced in dashed black line as a guide to the eye. The angular dispersion can be decreased to 0 with a sensitivity on the angular position of the prism as high as  $0.5\ \text{mrad}$ , corresponding to an angular dispersion of  $0.7\ \text{mrad nm}^{-1}$ . The general symmetry observed between positive and negative values of angular dispersion is due to the mapping of the different wavelengths in increasing or decreasing order. We also note that a slight asymmetry in the value of angular dispersion between positive and negative values of prism angle is due to the spectral dispersion imposed by the first prism.

Therefore, our procedure to accurately visualize the angular dispersion is as follows. Since angular dispersion



is proportional to the pulse front tilt  $p$ , we increase in a first step the temporal dispersion to make the inclination of the ellipse more apparent. We can then accurately remove the pulse front tilt. Finally, the prism insertion is adjusted to minimize the second order dispersion as shown for the signal framed in green ( $\phi_2 = 0$ ,  $p = 0$ ).

In more detail, our numerical simulations show that the misalignment from perfect parallelism of the prisms not only affects the angular dispersion, but also introduces slight second and third-order phases in the probe pulse. However, the variation of  $\phi_2$  over the range of prism misalignments (a line in Fig. 5) is typically of  $\pm 1000 \text{ fs}^2$ , reasonably smaller than the values of second order dispersion that we introduced using the prism insertion. The third-order phase is typically  $10^5 \text{ fs}^3$  and is apparent as a parabolic-shaped intensity pattern in the  $(t, k_y)$  traces.

We have numerically simulated the experimental cases A, B, C of Fig. 5. The analytical expression of the diffracted signal for a temporally chirped probe with higher-order phase is provided in "Methods". We used for the probe's parameters the second and third-order phases and the angular dispersion determined by the ZEMAX modeling of the misaligned compressor. The results of our simulations are shown as the rightmost column in Fig. 5. We find an excellent agreement between simulations and experiments.

The transient grating allows for detecting pulse front tilt in the direction perpendicular to the grating periodicity. We note that when we rotated the transient grating by

$90^\circ$ , i.e., along  $y$  direction, no variation in the arrival time was observable. This demonstrates that no element in our setup generates angular dispersion in the  $x$  direction, as one could expect (see Fig. 4). Therefore, the transient grating diagnostic is particularly helpful to accurately remove pulse front tilt even for faint changes in the deviation angle of the prism compressor.

## Discussion

We have developed an extremely localized in-situ diagnostic that allows a characterization and synchronization of a weak probe pulse with a higher intensity pump: the synchronization between pump and probe can be defined in a spatial domain of less than  $10 \mu\text{m}$  longitudinally inside the sample; we find the optimal point of probe pulse compression and we show the pulse front tilt can be removed. This diagnostic is highly flexible to very different pump–probe crossing geometries: because of the reconfigurability of the SLM, the pump can be directed to match any probe incidence angle—within the limits of the numerical apertures—to characterize the probe pulse. Then, the pump beam can be modified to another crossing geometry to perform imaging experiments. Therefore, this approach is extremely valuable for providing well-characterized ultrashort probe pulses for pump–probe imaging of ultrashort events under high magnification using single or multiple ultrashort probes over a wide spectrum and with different directions whatever the polarization. This approach can therefore be integrated into experimental setups for the study of



reproducible as well as non-reproducible phenomena, where each of the probe pulses can be individually neatly characterized<sup>1,17,19–21</sup>. Since it is a relatively fast measurement (less than 2 min at high resolution), it can be operated routinely to optimize compressors and to characterize the point of “zero-delay” (synchronization) between pump and probe pulses with an accuracy better than  $\sim 6$  fs (see Fig. 3). The determination of the absolute value of the zero-delay point requires using a transient grating with minimal length (typ. 10  $\mu\text{m}$ ) so as to best define the location of the synchronization. For the other diagnostics (pulse compression, pulse front tilt), increasing the extent of the transient grating is beneficial to increase the signal to noise ratio.

Our technique is valid for a wide range of pulse durations, as long as the pump can be independently characterized. We have routinely characterized probe pulses with durations up to  $\sim 300$  fs, for instance when the dispersion compensator is not present or misaligned. The micrometric size of the transient grating also allows for a very wide spectral bandwidth allowing very short probe pulses to be characterized. A possible limitation could occur for extremely short pump pulses (typ. below 30 fs), if the spectral extent cannot be neglected in the shaping by the SLM. In this case, spatio-temporal distortions could occur on the pump beam and dedicated strategies are anyway necessary. The blurring of the transient grating fringes is not a limitation in practice since this would require octave-spanning pulses.

The use of an SLM is highly beneficial since the convenient switch between arbitrary phase profiles makes it possible to combine our diagnostic with pump–probe techniques where structured beams are involved. These structured beams indeed have a number of applications, such as high aspect ratio micro-nano structuring<sup>58–62</sup>, laser welding<sup>63</sup> among many others, and opens up new perspectives for studying laser-dielectric interaction in the ultrafast regime<sup>3,64,65</sup>. The efficiency of SLMs is conventionally higher than 50% depending on the spatial frequency of the grating used. The maximal pulse energy can reach several mJ level and several 100's W average power<sup>66</sup>.

Our technique also remains valid even in presence of spherical aberration<sup>67</sup> since the interacting waves are collimated beams. Our technique could be usefully extended to detect potential pump–probe delay variations between probe angles as an example. This technique is, therefore, widely applicable to most of the ultrafast imaging and pump–probe experimental setups.

Our results have therefore a wide range of applicability, and we anticipate they will be particularly useful to characterize transient phenomena at micron-scale and laser-matter interaction within condensed matter.

## Methods

### SLM Phase mask

The phase mask applied to our SLM is split in two equal parts: the top half generates a plane-wave-like beam propagating at an angle  $\alpha - \theta$  and the lower half of the SLM mask symmetrically generates a beam propagating in direction  $\alpha + \theta$  toward the optical axis. The two beamlets cross at an angle of  $2\theta$ . We perform spatial filtering in the Fourier plane of the first lens ( $f = 750$  mm) after the SLM to select only the first diffraction order due to the SLM mask. The fringe period of the interfering two pump waves is  $\Lambda = \frac{\lambda_{\text{pump}}}{2n \sin \theta}$ . (For ease of reading, all angles are expressed in the dielectric medium of refractive index  $n$ , but wavelengths are expressed in vacuum)

### Bragg angle

The rotation angle  $\alpha$  applied to match the Bragg incidence condition for the probe pulse, is determined by  $\lambda_{\text{probe}} = 2n\Lambda \sin \alpha$ , such that:  $\sin(\alpha) = \sin(\theta) \frac{\lambda_{\text{probe}}}{\lambda_{\text{pump}}}$ . In this expression, we have neglected the variation of refractive index with wavelength as this modifies the angle by only a few percents.

### Diffraction efficiency

The diffraction efficiency of the probe pulse on the pump-induced grating can be derived using coupled-wave theory describing thick gratings, since the grating is much longer than its period  $\Lambda$  (see Suppl. Materials Fig. S1). We use in the following the work by Kogelnik which provides analytically the effect of detuning<sup>68</sup>. We note that identical results could be obtained using Four-Wave Mixing (FWM) theory. The samples investigated here (fused silica and sapphire) possess a large bandgap. This ensures that resonant 3-photon absorption is negligible and therefore that the nonlinear Kerr response is instantaneous<sup>69</sup>. We note that for  $\text{TiO}_2$  and smaller bandgap dielectrics, the  $\sim 6$  fs retardance observed in reference<sup>69</sup> is close to negligible in comparison with the pulse durations used here.

The coupled wave theory allows for deriving the signal diffracted in the first order of diffraction. The assumptions are that the incident beam and diffracted one are monochromatic plane waves, incident on an infinitely wide grating of thickness  $d$ . Those conditions are reasonably fulfilled in our experiments. In this framework, the diffraction efficiency, i.e., the ratio between the diffracted intensity in the first order at the exit of the grating  $|A_1(d)|^2$  and the incident beam intensity  $|A_i|^2$ , can be expressed as<sup>68</sup>:

$$\eta(\xi, \phi) = \frac{\sin^2 \sqrt{\xi^2 + \phi^2}}{1 + \left(\frac{\xi^2}{\phi^2}\right)} \quad (4)$$

where

$$\phi = \frac{\pi}{\lambda(1 - 2\sin^2\alpha)^{\frac{1}{2}}} d\delta n \quad (5)$$

$$\xi = \frac{2\pi n \sin^2\alpha}{\lambda^2(1 - 2\sin^2\alpha)} \delta\lambda \quad (6)$$

$\lambda$  is the probe central wavelength.  $\xi$  expresses the detuning, i.e., expresses how the diffraction efficiency reduces when the probe wavelength differs from the central wavelength at which the Bragg incidence is met. The wavelength detuning is written  $\delta\lambda$ . A similar relationship can also be expressed for the illumination angle detuning<sup>68</sup>.

Therefore, in our experimental conditions, where  $d \sim 30 \mu\text{m}$ ,  $\delta n \simeq 10^{-4}$  and  $\lambda = 0.4 \mu\text{m}$ ,  $\alpha = 6^\circ$  the diffraction efficiency is on the order of  $10^{-3}$ , which varies over a 30 fs probe pulse spectrum by less than 1%. The diffraction efficiency peak width exceeds 50 nm FWHM, i.e., much larger than the bandwidth of a 10 fs pulse centered at 395 nm.

Since the wavelength detuning is negligible over the probe pulse spectrum, the diffraction efficiency becomes:

$$\eta = \frac{|A_1(d)|^2}{|A_i|^2} = \sin^2\left(\pi \frac{d\delta n}{\lambda |\cos\alpha|}\right) \quad (7)$$

Then we can express the time-integrated diffracted intensity, with  $\tau$  being the delay of the probe with regard to the pump pulse:

$$\Sigma(\tau) = \int I_{\text{probe}}^{\text{1st order}}(t) dt \quad (8)$$

$$= \int \sin^2\left(\pi \frac{dn_2 I_{\text{pump}}(t)}{\lambda \cos\alpha}\right) I_{\text{probe}}(t - \tau) dt \quad (9)$$

$$\propto \left(\frac{n_2}{\lambda \cos\alpha}\right)^2 \int I_{\text{pump}}^2(t) I_{\text{probe}}(t - \tau) dt \quad (10)$$

provided that  $d\delta n \ll \lambda \cos\alpha$  which is fulfilled in our experimental conditions. Hence, the diffracted signal is proportional to the correlation function between  $I_{\text{pump}}^2$  and  $I_{\text{probe}}$ , as in TG-XFROG<sup>32,70</sup>.

### Setup

The Ti:Sapphire chirped pulse amplifier (CPA) laser source (Coherent Legend) delivers  $\sim 50$  fs pulses at 790 nm central wavelength and repetition rate 1 kHz. We split the beam in a pump and a probe, the latter is frequency doubled with a 50  $\mu\text{m}$  thick  $\beta$  barium borate (BBO) crystal. The pump pulse is then spectrally filtered to reduce its bandwidth to 12 nm FWHM, avoiding chromatic dispersion in

the beam shaping stage. The spectral transmission curve of the filter is nearly Gaussian so as to ensure the absence of pre-/post pulses. We spatially shape the pump beam using a spatial light modulator (SLM) in near-normal incidence. The input beam is expanded to quasi-uniformly illuminate the full active area of the SLM. We de-magnify the resulting shaped beam by a factor 208 using a 2f-2f arrangement using a first lens of focal length  $f_1 = 750$  mm and a second of focal length  $f_2 = 3.6$  mm (Microscope Objective Olympus MPLFLN  $\times 50$  with Numerical Aperture (NA) 0.8). Spatial filtering of the first order of diffraction is performed in the Fourier plane of the first lens<sup>71</sup> (not shown in the figure).

We pre-compensate the linear dispersion of the 395 nm probe with a folded prism compressor<sup>36</sup>. The probe beam is then de-magnified by a  $\times 1/278$  2f-2f arrangement (lens  $f_3 = 1000$  mm and the same  $\times 50$  microscope objective) so that the probe beam in the sample has a waist of 12  $\mu\text{m}$  (Rayleigh range of 1.1 mm) at the focus of the microscope objective.

Pump and probe energies and polarization states are controlled by half-waveplates and polarized beam splitters. Polarizations are controlled using independent half-waveplates placed on each beam path. The relative pump-probe delay is controlled with a motorized delay line in the probe beam, with a resolution of 3.3 fs. Figure S1 in the Supplementary Materials shows the pump and probe beam experimental characterizations.

After interaction in the sample, we collect the pulses with a second  $\times 50$  (N.A. 0.8) microscope objective (MO), as shown in Fig. 1d. When recording the diffracted signal from the transient grating, we filter out the residual pump signal. A relay lens images the back focal plane of the second MO on a CCD camera, which consequently records the far-field of the diffracted beams, which spatially separates the different orders of diffraction. We select out the +1 order by summing the signal over a region of  $\sim 100 \times 100$  pixels on the CCD. The measurement is performed in multishot regime, with a 14 bits camera in free-run mode, with illumination time chosen in the range 10 to 50 ms so as to use the whole dynamical range of the camera over which the response is linear.

### Pulse duration characterization

Before starting our experiments, we characterized the 790 nm pump pulse by self-referencing spectral interferometry (Wizzler<sup>TM</sup>) before the first microscope objective to measure and fully characterize in amplitude and phase the shortest pulse achievable. The pulse was compressed before the focusing microscope objective. Taking into account the dispersion of the microscope objective of 2300 fs<sup>2</sup> at 790 nm central wavelength (separately characterized), the pump pulse duration is 100 fs at the sample site. The probe pulse duration was measured

**Table 1** Parameters used in the experiments shown in the figures above: material, pump energy  $E_{\text{pump}}$  and half-crossing angle of the pump beams  $\theta$ .

	Fig. 2	Fig. 5	Fig. 3
Material	Schott D263	Sapphire	Sapphire
$E_{\text{pump}}$ ( $\mu\text{J}$ )	0.05–0.5	0.1	0.1
$\theta$ ( $^\circ$ )	14.4	12	12

In all cases, to match the Bragg incidence angle,  $\alpha$  is half of  $\theta$  because the probe central wavelength is half of the pump one.

before the prism compressor using an autocorrelator (APE pulseCheck<sup>TM</sup>).

### Shift in delay of the cross-correlation curve

When the sample is shifted upstream by  $d$ , the crossing point of the two pump beams is shifted downstream by a distance  $\Delta = |d|(n_g^{790} - 1)$ , at first order in the small angle. Then, Eq. (2) can be retrieved by considering the difference in optical paths between pump and probe from the initial case, again at first order in the small angles of the waves with the optical axis.

### Prism deviation measurement

The relative angle of rotation of the second prism in the prism compressor has been monitored by measuring the deviation angle of the reflection of a laser pointer onto one facet of the prism. The precision of this measurement was 0.3 mrad.

### Parameters for the experiments shown in the figures

Table 1 provides the main parameters of the experiments. All angles are given in material and correspond to a single grating pitch  $\Lambda = 1.1 \mu\text{m}$ . Difference on angle values are related to respective material indices.

### Numerical simulations of the misaligned prism compressor

The prism compressor and the 2f-2f optical arrangement were numerically simulated using ZEMAX software. This software enables a complete ray tracing over complex imaging systems and computes the optical path length. Hence it is then possible to model the dispersion effects up to the third order, including pulse front tilt and higher order dispersion. We note that the precise design of the Olympus MPLFLN  $\times 50$  microscope objective that we used experimentally is not available within this software. We replaced it by a  $\times 60$  microscope objective in our simulations. Since the dispersions are not precisely identical, we have adapted the dimensions of the folded prism compressor in the numerical simulations to precisely compensate the second order phase of the microscope objective and the lens. The parameters of the folded prism compressor were the following: the prisms are SF10

prisms with an apex angle of  $60^\circ$  degrees and a distance of 158 mm between the prisms. However, we note that the results obtained using this simulation are in excellent agreement with the experiments. The values of the higher order phases found using ZEMAX and used to simulate the traces are respectively: (A)  $\phi_2 = 1.1 \times 10^4 \text{ fs}^2$  and  $\phi_3 = -1.8 \times 10^5 \text{ fs}^3$ , (B)  $\phi_2 = 1.1 \times 10^4 \text{ fs}^2$  and  $\phi_3 = -1.6 \times 10^5 \text{ fs}^3$ , (C)  $\phi_2 = 3.1 \times 10^3 \text{ fs}^2$  and  $\phi_3 = -1.7 \times 10^5 \text{ fs}^3$ .

### Simulations of the diffracted signal in Fig. 5

The simulations of the diffracted signal in Fig. 5 are based on numerical integration of the following expression:

$$\begin{aligned} \Sigma(k_y, \tau) &= \int |\hat{A}^{+1}(k_y, t)|^2 dt \\ &\sim \iint |I_{\text{pump}} \cdot A_{\text{probe}}(y, t)|^2 e^{ik_y y} dt dy \end{aligned}$$

where  $\hat{A}^{+1}$  is the spatial Fourier transform of the diffracted amplitude, and  $A_{\text{probe}}$  is the amplitude of the probe pulse.  $\tilde{A}_{\text{probe}}(y, \omega) = E_0 e^{-T^2 \omega^2 / 4} e^{-y^2 / w_0^2} e^{-ip\omega y} e^{i(\frac{1}{2}\phi_2 \omega^2 + \frac{1}{6}\phi_3 \omega^3 + \dots)}$ , following the model by Akturk et al.<sup>57</sup>

The input pump pulse is modeled by a Gaussian pulse with 100 fs Full Width at Half-Maximum (FWHM). The probe is a 60 fs pulse with a small bump in the amplitude spectrum peaking at 390 nm so as to reproduce the experimental spectrum.

### Acknowledgements

Technical assistance by L. Furfaro, C. Billet, M. Dordor, and B. Guichardaz as well as fruitful discussions with D. Brunner, E. Lantz, and J.M. Dudley from FEMTO-ST (Besançon, France) are gratefully acknowledged. The authors acknowledge the financial supports of: European Research Council (ERC) 682032-PULSAR, Region Franche-Comte council (support to FRILIGHT platform), Labex ACTION ANR-11-LABX-0001-01; French RENATECH network, the EIPHI Graduate School ANR-17-EURE-0002 and I-SITE BFC project ANR-15-IDEX-0003.

### Author contributions

R.M., R.G., and F.C. developed the setup on an initial concept by C.X. and F.C. C.X. built the experimental interface for data acquisition, R.M. built the experimental setup, C.X. and F.C. performed the numerical simulations of beam propagation and diffracted signal. L.F. performed the numerical analysis of the pulse front tilt and dispersion using ZEMAX. R.M., R.G., and F.C. analyzed the experimental data. C.X. and F.C. made the analytical derivations. R.M. designed the figures, together with C.X. for Fig. 1. The manuscript was jointly written by C.X., R.M., and F.C. and revised by all authors.

### Conflict of interest

The authors declare no competing interests.

**Supplementary information** The online version contains supplementary material available at <https://doi.org/10.1038/s41377-021-00562-1>.

Received: 13 November 2020 Revised: 3 May 2021 Accepted: 26 May 2021  
Published online: 16 June 2021

### References

1. Matlis, N. H. et al. Snapshots of laser wakefields. *Nat. Phys.* **2**, 749–753 (2006).
2. Winkler, T. et al. Laser amplification in excited dielectrics. *Nat. Phys.* **14**, 74–79 (2018).

3. Hayasaki, Y. et al. Two-color pump-probe interferometry of ultra-fast light-matter interaction. *Sci. Rep.* **7**, 10405 (2017).
4. Clerici, M. et al. Wavelength scaling of terahertz generation by gas ionization. *Phys. Rev. Lett.* **110**, 253901 (2013).
5. Zhang, Z. L. et al. Manipulation of polarizations for broadband terahertz waves emitted from laser plasma filaments. *Nat. Photonics* **12**, 554–559 (2018).
6. Ghimire, S. & Reis, D. A. High-harmonic generation from solids. *Nat. Phys.* **15**, 10–16 (2019).
7. Vaillonis, A. et al. Evidence of superdense aluminium synthesized by ultrafast microexplosion. *Nat. Commun.* **2**, 445 (2011).
8. Li, Z.-Z. et al. O-FIB: far-field-induced near-field breakdown for direct nanowriting in an atmospheric environment. *Light Sci. Appl.* **9**, 41 (2020).
9. Pan, C. J. et al. Ultrafast optical response and ablation mechanisms of molybdenum disulfide under intense femtosecond laser irradiation. *Light Sci. Appl.* **9**, 80 (2020).
10. Downer, M. C., Fork, R. L. & Shank, C. V. Femtosecond imaging of melting and evaporation at a photoexcited silicon surface. *J. Opt. Soc. Am. B* **2**, 595–599 (1985).
11. von der Linde, D., Sokolowski-Tinten, K. & Bialkowski, J. Laser–solid interaction in the femtosecond time regime. *Appl. Surf. Sci.* **109–110**, 1–10 (1997).
12. Steinmeyer, G. et al. Frontiers in ultrashort pulse generation: pushing the limits in linear and nonlinear optics. *Science* **286**, 1507–1512 (1999).
13. Hentschel, M. et al. Attosecond metrology. *Nature* **414**, 509–513 (2001).
14. Niikura, H. et al. Probing molecular dynamics with attosecond resolution using correlated wave packet pairs. *Nature* **421**, 826–829 (2003).
15. Baum, P. & Zewail, A. H. Attosecond electron pulses for 4d diffraction and microscopy. *Proc. Natl Acad. Sci. USA* **104**, 18409–18414 (2007).
16. Goda, K., Tsia, K. K. & Jalali, B. Serial time-encoded amplified imaging for real-time observation of fast dynamic phenomena. *Nature* **458**, 1145–1149 (2009).
17. Nakagawa, K. et al. Sequentially timed all-optical mapping photography (STAMP). *Nat. Photonics* **8**, 695–700 (2014).
18. Gao, L. et al. Single-shot compressed ultrafast photography at one hundred billion frames per second. *Nature* **516**, 74–77 (2014).
19. Li, Z. Y. et al. Single-shot tomographic movies of evolving light-velocity objects. *Nat. Commun.* **5**, 3085 (2014).
20. Mikami, H., Gao, L. & Goda, K. Ultrafast optical imaging technology: principles and applications of emerging methods. *Nanophotonics* **5**, 497–509 (2016).
21. Ehn, A. et al. FRAME: femtosecond videography for atomic and molecular dynamics. *Light Sci. Appl.* **6**, e17045–e17045 (2017).
22. Wang, P., Liang, J. Y. & Wang, L. V. Single-shot ultrafast imaging attaining 70 trillion frames per second. *Nat. Commun.* **11**, 2091 (2020).
23. Qi, D. L. et al. Single-shot compressed ultrafast photography: a review. *Adv. Photonics* **2**, 014003 (2020).
24. Kim, T. et al. Picosecond-resolution phase-sensitive imaging of transparent objects in a single shot. *Sci. Adv.* **6**, eaay6200 (2020).
25. Hernandez-Rueda, J., Witcher, J. J. & Krol, D. M. Ultrafast time-resolved spectroscopy of a fs-laser-induced plasma inside glass using a super-continuum probe beam. *Appl. Phys. A* **125**, 591 (2019).
26. Duguay, M. A. & Hansen, J. W. Direct measurement of picosecond lifetimes. *Opt. Commun.* **1**, 254–256 (1969).
27. Sala, K. & Richardson, M. C. Optical Kerr effect induced by ultrashort laser pulses. *Phys. Rev. A* **12**, 1036–1047 (1975).
28. Ho, P. P., Yu, W. & Alfano, R. R. Relaxation of the optical Kerr effect of anisotropic molecules in mixed liquids. *Chem. Phys. Lett.* **37**, 91–96 (1976).
29. Yu, B. L. et al. Femtosecond optical Kerr shutter using lead–bismuth–gallium oxide glass. *Opt. Commun.* **215**, 407–411 (2003).
30. Kane, D. J. & Trebino, R. Single-shot measurement of the intensity and phase of an arbitrary ultrashort pulse by using frequency-resolved optical gating. *Opt. Lett.* **18**, 823–825 (1993).
31. Lebugle, M. et al. High temporal resolution and calibration in pump–probe experiments characterizing femtosecond laser–dielectrics interaction. *Appl. Phys. A* **120**, 455–461 (2015).
32. Lee, D., Gabolde, P. & Trebino, R. Toward single-shot measurement of a broadband ultrafast continuum. *J. Opt. Soc. Am. B* **25**, A34–A40 (2008).
33. Ermolov, A. et al. Characterization of few-fs deep-UV dispersive waves by ultra-broadband transient-grating XFROG. *Opt. Lett.* **41**, 5535–5538 (2016).
34. Sarpe, C. et al. Real-time observation of transient electron density in water irradiated with tailored femtosecond laser pulses. *New J. Phys.* **14**, 075021 (2012).
35. Guild, J. B., Xu, C. & Webb, W. W. Measurement of group delay dispersion of high numerical aperture objective lenses using two-photon excited fluorescence. *App. Opt.* **36**, 397–401 (1997).
36. Fork, R. L., Martinez, O. E. & Gordon, J. P. Negative dispersion using pairs of prisms. *Opt. Lett.* **9**, 150–152 (1984).
37. Chauhan, V., Cohen, J. & Trebino, R. Simple dispersion law for arbitrary sequences of dispersive optics. *Appl. Opt.* **49**, 6840–6844 (2010).
38. Sacks, Z., Mourou, G. & Danielius, R. Adjusting pulse-front tilt and pulse duration by use of a single-shot autocorrelator. *Opt. Lett.* **26**, 462–464 (2001).
39. Akturk, S. et al. Measuring pulse-front tilt in ultrashort pulses using GRENOUILLE. *Opt. Express* **11**, 491–501 (2003).
40. Grunwald, R. et al. Ultrashort-pulse wave-front autocorrelation. *Opt. Lett.* **28**, 2399–2401 (2003).
41. Bock, M. et al. Reconfigurable wavefront sensor for ultrashort pulses. *Opt. Lett.* **37**, 1154–1156 (2012).
42. Guang, Z. et al. Complete characterization of a spatiotemporally complex pulse by an improved single-frame pulse-measurement technique. *J. Opt. Soc. Am. B* **31**, 2736–2743 (2014).
43. Li, Z. Y., Miyayaga, N. & Kawanaka, J. Single-shot real-time detection technique for pulse-front tilt and curvature of femtosecond pulsed beams with multiple-slit spatiotemporal interferometry. *Opt. Lett.* **43**, 3156–3159 (2018).
44. Eichler, H. J., Enterlein, G. & Langhans, D. Investigation of the spatial coherence of a laser beam by a laser-induced grating method. *Appl. Phys.* **23**, 299–302 (1980).
45. Eichler, H. J., Klein, U. & Langhans, D. Coherence time measurement of picosecond pulses by a light-induced grating method. *Appl. Phys.* **21**, 215–219 (1980).
46. Eichler, H. J., Günter, P. & Pohl, D. W. *Laser-Induced Dynamic Gratings* (Springer, 1986).
47. Högemann, C., Pauchard, M. & Vauthey, E. Picosecond transient grating spectroscopy: The nature of the diffracted spectrum. *Rev. Sci. Instrum.* **67**, 3449–3453 (1996).
48. Wörner, H. J. et al. Following a chemical reaction using high-harmonic interferometry. *Nature* **466**, 604–607 (2010).
49. Iaconis, C. & Walmsley, I. A. Self-referencing spectral interferometry for measuring ultrashort optical pulses. *IEEE J. Quantum Electron.* **35**, 501–509 (1999).
50. Moulet, A. et al. Single-shot, high-dynamic-range measurement of sub-15 fs pulses by self-referenced spectral interferometry. *Opt. Lett.* **35**, 3856–3858 (2010).
51. DeLong, K. W., Trebino, R. & Kane, D. J. Comparison of ultrashort-pulse frequency-resolved-optical-gating traces for three common beam geometries. *J. Opt. Soc. Am. B* **11**, 1595–1608 (1994).
52. Major, A. et al. Dispersion of the nonlinear refractive index in sapphire. *Opt. Lett.* **29**, 602–604 (2004).
53. Santran, S. et al. Precise and absolute measurements of the complex third-order optical susceptibility. *J. Opt. Soc. Am. B* **21**, 2180–2190 (2004).
54. Lorient, V. et al. Measurement of high order Kerr refractive index of major air components. *Opt. Express* **17**, 13429–13434 (2009).
55. Boyd, R. *Nonlinear Optics*. 3rd edn (Academic Press, Amsterdam, 2008).
56. Query, M. R. *Optical constants*. Tech. Rep. (Missouri Univ-Kansas City, 1985).
57. Akturk, S. et al. Pulse-front tilt caused by spatial and temporal chirp. *Opt. Express* **12**, 4399–4410 (2004).
58. Gattass, R. R. & Mazur, E. Femtosecond laser micromachining in transparent materials. *Nat. Photonics* **2**, 219–225 (2008).
59. Bellouard, Y. et al. Fabrication of high-aspect ratio, micro-fluidic channels and tunnels using femtosecond laser pulses and chemical etching. *Opt. Express* **12**, 2120–2129 (2004).
60. Bhuyan, M. K. et al. High aspect ratio nanochannel machining using single shot femtosecond Bessel beams. *Appl. Phys. Lett.* **97**, 081102 (2010).
61. Courvoisier, F., Stoian, R. & Couairon, A. [invited] ultrafast laser micro- and nano-processing with nondiffracting and curved beams: invited paper for the section: hot topics in ultrafast lasers. *Opt. Laser Technol.* **80**, 125–137 (2016).
62. Stoian, R. et al. Ultrafast Bessel beams: advanced tools for laser materials processing. *Adv. Opt. Technol.* **7**, 165–174 (2018).
63. Watanabe, W. et al. Space-selective laser joining of dissimilar transparent materials using femtosecond laser pulses. *Appl. Phys. Lett.* **89**, 021106 (2006).
64. Hayasaki, Y. et al. Photo-acoustic sub-micrometer modifications of glass by pair of femtosecond laser pulses. *Opt. Mater. Express* **2**, 691–699 (2012).
65. Bergner, K. et al. Time-resolved tomography of ultrafast laser-matter interaction. *Opt. Express* **26**, 2873–2883 (2018).

66. Zhu, G. et al. Investigation of the thermal and optical performance of a spatial light modulator with high average power picosecond laser exposure for materials processing applications. *J. Phys. D: Appl. Phys.* **51**, 095603 (2018).
67. Booth, M. J., Neil, M. A. A. & Wilson, T. Aberration correction for confocal imaging in refractive-index-mismatched media. *J. Microsc.* **192**, 90–98 (1998).
68. Kogelnik, H. Coupled wave theory for thick hologram gratings. *Bell Syst. Tech.* **48**, 2909–2947 (1969).
69. Hofmann, M. et al. Noninstantaneous polarization dynamics in dielectric media. *Optica* **2**, 151–157 (2015).
70. Sweetser, J. N., Fittinghoff, D. N. & Trebino, R. Transient-grating frequency-resolved optical gating. *Opt. Lett.* **22**, 519–521 (1997).
71. Froehly, L. et al. Spatiotemporal structure of femtosecond Bessel beams from spatial light modulators. *J. Opt. Soc. Am. A, Opt. Image Sci.* **31**, 790–793 (2014).